


F

Repeatability					
Your "r" performance					
	Fat route	Protein route	Lactose route	Urea route	BHB
	g/100g	nitrogen g/100g *	g/100g	mg/dl	mmol/L
Sample 1	0.000	0.000	0.00	0.40	
Sample 2	0.000	0.010	0.00	0.20	
Sample 3	0.010	0.000	0.00	0.70	
Sample 4	0.000	0.000	0.00	0.80	
Sample 5	0.000	0.000	0.00	0.10	
Sample 6	0.000	0.000	0.00	0.60	
Sample 7	0.000	0.000	0.00	0.70	
Sample 8	0.000	0.000	0.00	0.80	
Sample 9	0.000	0.000	0.00	0.60	
Sample 10	0.000	0.000	0.00	0.80	
If the repeatability is smaller than the limit the cell is in green if there is a sample with a "r" bigger than the limit the cell is in red. Please check table II in correspondence of the parameter and your lab code.					
Limits					
	Fat route	Protein route	Lactose route	Urea route	BHB
	g/100g	nitrogen g/100g *	g/100g	mg/dl	mmol/L
	ISO 1211 IDF 1D	ISO 8968 IDF 20	ISO 22662 IDF 198	ISO 14637 IDF 195	Indicative
	<= 0.043	<= 0.038	<= 0.06	<= 2.58	<= 0.04


Your Z-Score PT					
	Fat route	Protein route	Lactose route	Urea route	BHB
Sample 1	0.04	-0.41	-1.53	0.04	
Sample 2	-0.17	0.16	-1.57	-0.14	
Sample 3	-0.05	-0.40	-1.65	0.36	
Sample 4	-0.14	0.09	-1.42	-0.52	
Sample 5	0.12	-0.20	-1.68	-0.27	
Sample 6	-0.10	0.09	-1.53	-0.06	
Sample 7	-0.12	-0.22	-1.64	-0.74	
Sample 8	-0.11	-0.05	-1.75	0.00	
Sample 9	-0.94	-0.53	-1.75	0.21	
Sample 10	0.00	-0.47	-1.58	-0.02	
Interpretation Z-Score					
Z-Score < -3	-3 <= Z-Score < -2	-2 <= Z-Score <= 2	2 < Z-Score <= 3	Z-Score > 3	
Poor	Moderate	Good	Moderate	Poor	

ZSCORE-PT - FAT routine




Part. code	Round	SP01	SP02	SP03	SP04	SP05	SP06	SP07	SP08	SP09	SP10	Yellow	Red	White
27	RT0323	0.0	-0.2	-0.1	-0.1	0.1	-0.1	-0.1	-0.1	-0.9	0.0	0%	0%	100%
25	RT0322	-0.5	-0.1	0.2	0.5	0.5	0.3	0.4	-1.3	0.5	-0.7	0%	0%	100%
22	RT0321	0.6	0.5	0.6	0.5	0.1	0.5	-0.4	0.3	0.2	0.6	0%	0%	100%
12	RT0320	0.6	0.3	0.6	0.6	0.8	0.6	-0.7	0.2	0.6	0.3	0%	0%	100%
8	RT0919	3.9	2.6	2.9	1.6	1.4	2.4	2.8	2.2	1.8	1.0	50%	10%	40%
12	RT0918	0.3	0.1	-0.2	0.1	-0.7	0.0	-0.3	0.1	0.4	-0.4	0%	0%	100%
16	RT0916	-0.5	-0.2	-0.5	-0.3	-0.3	-1.0	-0.7	-0.2	-0.3	-0.3	0%	0%	100%

ZSCORE-PT - PROTEIN routine




Part. code	Round	SP01	SP02	SP03	SP04	SP05	SP06	SP07	SP08	SP09	SP10	Yellow	Red	White
21	RT0323	-0.4	0.2	-0.4	0.1	-0.2	0.1	-0.2	0.0	-0.5	-0.5	0%	0%	100%
13	RT0322	0.9	1.0	0.4	0.2	0.0	0.9	0.0	1.1	0.2	1.3	0%	0%	100%
22	RT0321	-0.6	-0.6	-0.8	-0.3	-0.2	-0.5	-0.7	-0.7	-0.3	-0.4	0%	0%	100%
12	RT0320	1.1	1.0	1.5	1.3	1.4	1.5	1.3	0.8	2.1	0.2	10%	0%	90%
8	RT0919	0.3	0.6	0.6	0.7	0.4	0.4	0.6	0.9	0.5	0.6	0%	0%	100%
12	RT0918	0.0	1.3	1.5	0.7	1.8	1.0	1.9	1.3	0.7	1.5	0%	0%	100%
16	RT0916	-0.3	0.0	-0.3	-0.5	-0.4	-0.6	0.0	-0.3	-0.7	-0.8	0%	0%	100%

ZSCORE-PT - LACTOSE routine



Part. code	Round	SP01	SP02	SP03	SP04	SP05	SP06	SP07	SP08	SP09	SP10	Yellow	Red	White
25	RT0323	-1.5	-1.6	-1.6	-1.4	-1.7	-1.5	-1.6	-1.8	-1.8	-1.6	0%	0%	100%
24	RT0322	0.5	0.6	0.6	0.4	0.5	0.4	0.5	0.8	0.4	0.1	0%	0%	100%
20	RT0321	0.8	1.0	0.9	0.4	0.7	0.5	1.1	0.5	0.4	0.6	0%	0%	100%
11	RT0320	-0.6	-0.6	-0.7	-1.1	-1.1	-1.2	-1.1	-0.7	-0.8	-1.6	0%	0%	100%
8	RT0919	-0.8	-0.7	-1.0	-1.2	-1.3	-1.1	-1.4	-1.4	-1.4	-1.6	0%	0%	100%
10	RT0918	0.8	0.9	1.1	0.9	1.2	1.0	0.8	1.0	1.0	1.2	0%	0%	100%
14	RT0916	-0.3	0.0	-0.5	-0.4	-0.3	-0.5	-0.5	-0.2	-0.1	-0.5	0%	0%	100%

ZSCORE-PT - UREA routine



Part. code	Round	SP01	SP02	SP03	SP04	SP05	SP06	SP07	SP08	SP09	SP10	Yellow	Red	White
26	RT0323	0.0	-0.1	0.4	-0.5	-0.3	-0.1	-0.7	0.0	0.2	0.0	0%	0%	100%
25	RT0322	-0.5	-0.3	-0.6	-0.4	-0.5	-0.1	-0.8	-0.5	-0.5	-0.4	0%	0%	100%
20	RT0321	0.7	0.7	0.8	0.5	0.5	0.5	0.4	0.3	0.4	0.2	0%	0%	100%
12	RT0320	0.4	0.7	0.5	0.6	0.7	0.6	0.8	0.9	0.8	0.8	0%	0%	100%
7	RT0919	-0.4	-0.1	-0.6	-0.6	-0.3	-0.3	-0.7	-0.9	-0.5	-0.2	0%	0%	100%
9	RT0918	-0.7	-0.3	-0.2	-0.5	-0.6	-0.4	-0.8	-0.3	-0.2	-0.3	0%	0%	100%
13	RT0916	-0.4	-0.1	-0.1	0.0	0.3	0.2	0.3	0.2	0.2	0.4	0%	0%	100%

G	Your Z-Score PT					
		Fat _{ref}	Protein _{ref}	Lactose _{ref}	Urea _{ref}	SCC _{ref/alt}
	Sample 1					0.56
	Sample 2					0.70
	Sample 3					0.48
	Sample 4					0.53
	Sample 5					0.49
	Sample 6					0.47
	Sample 7					0.32
	Sample 8					0.52
	Sample 9					-0.12
	Sample 10					-0.13
	Your Z-Score Fix					
		Fat _{ref}	Protein _{ref}	Lactose _{ref}	Urea _{ref}	SCC _{ref/alt}
	Sample 1					1.06
	Sample 2					1.08
	Sample 3					0.81
	Sample 4					0.59
	Sample 5					0.82
	Sample 6					0.87
	Sample 7					0.57
	Sample 8					0.92
	Sample 9					1.38
	Sample 10					0.60
	If there is a sample with a "z-score" in the yellow or red area please check table VI and VII in correspondence of your lab code.					
	Interpretation Z-Score					
	Z-Score < -3	-3 ≤ Z-Score < -2	-2 ≤ Z-Score ≤ 2	2 < Z-Score ≤ 3	Z-Score > 3	
	Poor	Moderate	Good	Moderate	Poor	

